Perchloroethylene 10 to 80 µg/L

Order No. 81 01 551

Application Range		
Determination of perchloroethylene in water/waste water		
Dräger-Tube: Perchloroethylene 0.1/a		
Measuring range:	10 to 80 µg/L	
Number of Strokes (n): 8		
Typical Stroke Time:	2 to 3 minutes	
Measurement Time:	approx. 20 minutes	
Sample Volume:	ple Volume: 200 mL	
Color Change:	yellow white \rightarrow grey blue	
Temperature Range:	5 to 30 °C	
pH-Measurement: not necessary		

System Parameters

Measurement Range [µg/L]	Standard Deviation [%]	Temperature [°C]	Parar B	neters C
10 to 80	30	5 to 30	70	- 0.1

Evaluation of Measurement

Calculate perchloroethylene concentration:

$$Y_{[[\mu g/L]} = A \bullet B \bullet (X_{[ppm]} + C)$$

Cross Sensitivity

Dichloromethane, chlorobenzene, chloroform, 1,1-dichloroethane and 1,2-dichloroethane are indicated with lower sensitivity. Trichloroethylene is indicated with nearly the same sensitivity. Petroleum hydrocarbons, benzene, carbon tetrachloride, toluene, 1,1,1-trichloroethane and xylene are not indicated.



T-193-200

Chlorinated Hydrocarbons qualitative in oil Order No. 81 01 551

Application Range

Determination of volatile chlorinated hydrocarbons

in oil sludges/oil emulsions

5		
Dräger-Tube:	Perchloroethylene 0.1/a	
Measuring range:	ng range: qualitative	
Number of Strokes (n):	maximum 10	
Typical Stroke Time:	2 to 3 minutes	
Measurement Time:	approx. 2 to 20 minutes	
Sample Volume:	approx. 0.5 g	
Color Change:	yellow white \rightarrow grey blue	
Temperature Range:	10 to 25 °C	
pH-Measurement:	not necessary	

Information of Measurement

- Approx. 0.5 g oil sample has to be shaken intensively with 1 L de-ionized water for 2 minutes in a laboratory bottle.
- The solution must be filtered through an analysis funnel with a round filter (black ribbon) directly into the gas wash bottle up to the 200 mL mark.

Evaluation of Measurement

The measurement evaluation is qualitative (yes or no)

Cross Sensitivity

Chlorobenzene, 1,1-dichloroethane, 1,2-dichloroethane, dichloromethane, perchloroethylene, trichloroethylene and trichloromethane are indicated. Carbon tetrachloride and 1,1,1-trichloroethane are not indicated.



Chlorinated Hydrocarbons qualitative in multiple phase Order No. 81 01 551

Application Range

Determination of volatile chlorinated hydrocarbons in multiple phase

pilase		
Dräger-Tube:	e: Perchloroethylene 0.1/a	
Measuring range:	qualitative	
Number of Strokes (n):	maximum 10	
Typical Stroke Time:	2 to 3 minutes	
Measurement Time:	approx. 2 to 20 minutes	
Sample Volume:	200 mL	
Color Change:	yellow white → grey blue	
Temperature Range:	rature Range: 10 to 25 °C	
pH-Measurement: not necessary		

Information of Measurement

- Mix a multiple phase sample which consists of e.g. 250 g water, 10 g fixed phase and 10 g oil part (about 300 mL) is mixed with approx. 5 g activated coal. It must rest for 3 minutes and then be shaken for 1 min.
- 0.2 g hydrophobated peat is added and the it must be shaken for 1 minute.
- The liquid is filled into the gas wash bottle up to the 200 mL mark.

Evaluation of Measurement

The measurement evaluation is qualitative (yes or no)

Cross Sensitivity

Chlorobenzene, 1,1-dichloroethane, 1,2-dichloroethane, dichloromethane, perchloroethylene, trichloroethylene and trichloromethane are indicated. Carbon tetrachloride and 1,1,1-trichloroethane are not indicated.



Chlorinated Hydrocarbons qualitative in soil Order No. 81 01 551

Application Range

Determination of volatile chlorinated hydrocarbons in soil		
Dräger-Tube:	Perchloroethylene 0.1/a	
Measuring range:	g range: qualitative	
Number of Strokes (n):	maximum 10	
Typical Stroke Time:	Time: 2 to 3 minutes	
Measurement Time:	approx. 2 to 20 minutes	
Sample Volume:	me: 20 g	
Color Change:	yellow white → grey blue	
Temperature Range:10 to 25 °C		
pH-Measurement: not necessary		

Information of Measurement

- 20 g soil is suspended completely with 100 mL de-ionized water and 1 mL surfactant solutions (2 mass % Extran AP 13, Merck).
- The precipitate must rest for approx. 1 minute, until the particles have settled to the bottom; the liquid above the particles has to be filled into the wash bottle
- The remaining precipitate has to be shaken two times with 50 mL de-ionized water and the liquid above the particles has to be filled into the wash bottle
- The gas wash bottle is filled up with de-ionized water up to 200 mL mark.

Evaluation of Measurement

The measurement evaluation is qualitative (yes or no)

Cross Sensitivity

Chlorobenzene, 1,1-dichloroethane, 1,2-dichloroethane, dichloromethane, perchloroethylene, trichloroethylene and trichloromethane are indicated. Carbon tetrachloride and 1,1,1-trichloroethane are not indicated.



Trichloroethylene 10 to 100 µg/L

Order No. 81 01 551

Determination of trichloroethylene in water/waste water Dräger-Tube: Perchloroethylene 0.1/a Measuring range: 10 to 100 µg/L Number of Strokes (n): 4 Typical Stroke Time: 2 to 3 minutes Measurement Time: approx. 10 minutes Sample Volume: 200 mL Color Change: yellow white → grey blue Temperature Range: 5 to 30 °C pH-Measurement: not necessary	Application Range			
Measuring range:10 to 100 μg/LNumber of Strokes (n):4Typical Stroke Time:2 to 3 minutesMeasurement Time:approx. 10 minutesSample Volume:200 mLColor Change:yellow white → grey blueTemperature Range:5 to 30 °C	Determination of trichloroethylene in water/waste water			
Number of Strokes (n):4Typical Stroke Time:2 to 3 minutesMeasurement Time:approx. 10 minutesSample Volume:200 mLColor Change:yellow white → grey blueTemperature Range:5 to 30 °C	Dräger-Tube: Perchloroethylene 0.1/a			
Typical Stroke Time:2 to 3 minutesMeasurement Time:approx. 10 minutesSample Volume:200 mLColor Change:yellow white → grey blueTemperature Range:5 to 30 °C	Measuring range: 10 to 100 µg/L			
Measurement Time:approx. 10 minutesSample Volume:200 mLColor Change:yellow white → grey blueTemperature Range:5 to 30 °C	Number of Strokes (n):	4		
Sample Volume:200 mLColor Change:yellow white → grey blueTemperature Range:5 to 30 °C	Typical Stroke Time:	2 to 3 minutes		
Color Change:yellow white → grey blueTemperature Range:5 to 30 °C	Measurement Time: approx. 10 minutes			
Temperature Range:5 to 30 °C	Sample Volume:	200 mL		
·····p································	Color Change:	yellow white \rightarrow grey blue		
pH-Measurement: not necessary	Temperature Range:	5 to 30 °C		
	pH-Measurement: not necessary			

System Parameters

Measurement Range [µg/L]	Standard Deviation [%]	Temperatur [°C]	e Parameters B C
10 to 100	30	5 to 10 11 to 20 21 to 30	120 -0.01

Evaluation of Measurement

Calculate trichloroethylene concentration:

$$Y_{[\mu g/L]} = A \bullet B \bullet (X_{[ppm]} + C)$$

Cross Sensitivity

Dichloromethane, chlorobenzene, chloroform, 1,1-dichloroethane and 1,2-dichloroethane are indicated with lower sensitivity. Perchloroethylene is indicated with nearly the same sensitivity. Carbon tetrachloride and 1,1,1-trichloroethane are not indicated.

